

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite Mobile Speed Enforcement System

Customer: South Wales Police **Serial Number:** LW06290
Certificate Number: 8367
Date of Calibration: 02 October 2024 **Next Calibration:** 01 October 2025
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 20.000092 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 Horizontal - Pass
Eye Safety - Pass

Range \pm 0.5 ft
Actual 156.3 ft 356.6 ft 488.4 ft
Indicated 156.0 ft 356.5 ft 488.6 ft

Speed \pm 1 MPH
Actual 26 MPH -72 MPH -50 MPH 34 MPH 62 MPH
Indicated 26 MPH -72 MPH -50 MPH 34 MPH 62 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK



L. Korcz
Technical Signatory

TRUVELO (UK) LTD