

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer: Gwent Police **Serial Number:** LW06269
Certificate Number: 6342
Date of Calibration: 16 August 2022 **Next Calibration:** 15 August 2023
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 20.000120 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.0 ft	284.2 ft	488.5 ft

Speed ± 1 MPH

Actual	65 MPH	-23 MPH	-31 MPH	50 MPH	35 MPH
Indicated	65 MPH	-23 MPH	-31 MPH	50 MPH	35 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK



L K Orcz

Technical Signatory

TRUVELO (UK) LTD

Tel: 020 8847 4400

email: info@truelouk.com