## **CALIBRATION CERTIFICATE**

for

## **Kustom LASERwitness Lite**

**Mobile Speed Enforcement System** 

**Customer:** 

South Wales Police

Serial Number:

LW06289

Certificate Number:

8319

**Date of Calibration:** 

24 September 2024

**Next Calibration:** 

23 September 2025

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

**Calibration Results:** 

Checksum

- CC85

Oscillator Frequency

- 19,999827 MHz

Self Test Sequence

Vertical

Horizontal

- Pass

External Controls

1 400

- Pass

Aiming

- Pass

Eye Safety

PassPass

Range ± 0.5 ft Actual

156.3 ft

356.6 ft

488.4 ft

Indicated

156.3 ft

356.7 ft

488.5 ft

Speed ± 1 MPH

Actual

73 MPH

-25 MPH

-57 MPH

39 MPH

44 MPH

Indicated

**73 MPH** 

-25 MPH

-57 MPH

**39 MPH** 

44 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truyelo VK

Technical Signatory

TRUVELO (UK) LTD

Tel:020 8847 4400

email:info@truvelouk.com