

CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: Gwent Police **Serial Number:** LW06269
Date of Calibration: 27 August 2024 **Certificate Number:** 8268
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG **Next Calibration:** 26 August 2025

Calibration Results:

- Checksum - CC85
- Oscillator Frequency - 20.000098 MHz
- Self Test Sequence - Pass
- External Controls - Pass
- Aiming
 - Vertical - Pass
 - Horizontal - Pass
- Eye Safety - Pass

Range ± 0.5 ft

Actual	156.3 ft	356.6 ft	488.4 ft
Indicated	156.2 ft	356.4 ft	488.7 ft


Speed ± 1 MPH

Actual	22 MPH	-51 MPH	73 MPH	65 MPH	-39 MPH
Indicated	22 MPH	-51 MPH	73 MPH	65 MPH	-39 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 10 ft - 4500 ft, Speed 0 mph - 200 mph
The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo-UK



L. Kopez

Technical Signatory

TRUVELO (UK) LTD

TRU-CAL-CER-005

Tel: 020 8847 4400

email: info@truvelouk.com

