

# CALIBRATION CERTIFICATE

for

**Kustom LASERwitness Lite**

**Mobile Speed Enforcement System**

**Customer:** Gwent Police

**Serial Number:** LW06262

**Certificate Number:** 6219

**Date of Calibration:** 27 June 2022

**Next Calibration:** 26 June 2023

**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

## Calibration Results:

Checksum - CC85  
Oscillator Frequency - 19.999936 MHz  
Self Test Sequence - Pass  
External Controls - Pass  
Aiming - Vertical - Pass  
                  - Horizontal - Pass  
Eye Safety - Pass

## Range ± 0.5 ft

Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.2 ft	284.4 ft	488.7 ft

## Speed ± 1 MPH

Actual	74 MPH	26 MPH	-36 MPH	-42 MPH	63 MPH
Indicated	74 MPH	26 MPH	-36 MPH	-42 MPH	63 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

  
L Korcz

Technical Signatory

**TRUVELO (UK) LTD**