

# CALIBRATION CERTIFICATE

for  
**Kustom LASERwitness Lite**  
Mobile Speed Enforcement System

**Customer:** South Wales Police                      **Serial Number:** LW06283  
**Certificate Number:** 6069  
**Date of Calibration:** 13 May 2022                      **Next Calibration:** 12 May 2023  
**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

**Calibration Results:**

Checksum - CC85  
Oscillator Frequency - 19.999924 MHz  
Self Test Sequence - Pass  
External Controls - Pass  
Aiming - Vertical - Pass  
              - Horizontal - Pass  
Eye Safety - Pass

Range ± 0.5 ft  
Actual                      156.3 ft                      284.5 ft                      488.4 ft  
Indicated                      156.2 ft                      284.4 ft                      488.8 ft

Speed ± 1 MPH  
Actual                      25 MPH                      -51 MPH                      -73 MPH                      62 MPH                      37 MPH  
Indicated                      25 MPH                      -51 MPH                      -73 MPH                      62 MPH                      37 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK  
  
L Korcz  
Technical Signatory

---

**TRUVELO (UK) LTD**

---