CALIBRATION CERTIFICATE

for

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:

South Wales Police

Serial Number:

LW06299

Certificate Number:

7871

Date of Calibration:

07 March 2024

Next Calibration:

06 March 2025

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum

- CC85

Oscillator Frequency

- 19.999693 MHz

Self Test Sequence

- Pass

External Controls

- Pass

576 767

Vertical - Pass

Aiming

Horizontal - Pass

Eye Safety

- Pass

Range ± 0.5 ft

Actual

156.3 ft

356.6 ft

488.4 ft

Indicated

156.3 ft

356.7 ft

488.7 ft

Speed ± 1 MPH

Actual

27 MPH

-38 MPH

50 MPH

60 MPH

-72 MPH

Indicated

27 MPH

-38 MPH

50 MPH

60 MPH

-72 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

L Kørcz Technidal Signatory

TRUVELO (UK) LTD

Tel:020 8847 4400

email:info@truvelouk.com