

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: South Wales Police **Serial Number:** LW06290

Certificate Number: 6387

Date of Calibration: 08 September 2022 **Next Calibration:** 07 September 2023

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 20.000088 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming Vertical - Pass
 Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft
Actual 156.3 ft 284.5 ft 488.4 ft
Indicated 155.9 ft 284.2 ft 488.4 ft

Speed ± 1 MPH
Actual 33 MPH 72 MPH -43 MPH -26 MPH 59 MPH
Indicated 33 MPH 72 MPH -43 MPH -26 MPH 59 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK



L Korcz
Technical Signatory

TRUVELO (UK) LTD