

# CALIBRATION CERTIFICATE

for

**Kustom LASERwitness Lite**

**Mobile Speed Enforcement System**

**Customer:** South Wales Police **Serial Number:** LW06287

**Certificate Number:** 6465

**Date of Calibration:** 11 October 2022 **Next Calibration:** 10 October 2023

**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

## Calibration Results:

Checksum - CC85  
Oscillator Frequency - 20.000027 MHz  
Self Test Sequence - Pass  
External Controls - Pass  
Aiming - Vertical - Pass  
Horizontal - Pass  
Eye Safety - Pass

Range ± 0.5 ft			
Actual	156.3 ft	284.5 ft	488.4 ft
Indicated	156.3 ft	284.5 ft	488.6 ft

Speed ± 1 MPH					
Actual	38 MPH	74 MPH	-26 MPH	-55 MPH	62 MPH
Indicated	38 MPH	74 MPH	-26 MPH	-55 MPH	62 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz  
Technical Signatory

**TRUVELO (UK) LTD**